

GR265 G/E (Gas / Electric Vaccine Storage Refrigerator & Ice Pack Freezer)



WHO Test Procedure	E3/PROC/5
WHO Specifications	E3/RF2



Specifications

Vaccine Storage Capacity (L)	Refrigerator	57
Frozen Ice Pack Storage Capacity (L)	Freezer	9.6
Manufacturers Gross Volume (L)	Refrigerator	180
	Freezer	32

Type Of Kerosene Burner/Gas Jet Size	Jet Size 13
--------------------------------------	-------------

Hot Zone Appliance (+43°C)	YES
Temperate Zone Appliance (+32°C)	YES

Power Consumption / 24hr period (approx)***

LPG (g)	+32°C	660g
	+43°C	760g
Electric (kWh)	+32°C	6
	+43°C	8
Kerosene	N/A	

Performance Information

Hold Over Time During Power Cut	5.5 hours(+32°C)	
	1.5 hours (+43°C)	
Ice pack Freezing with Vaccine	+32°C	4.2kg / 24hrs
	+43°C	1.8kg / 24hrs
Ice pack Freezing without Vaccine	+32°C	N/A
	+43°C	N/A

Dimensions

Boxed Dimensions (mm)	(H)1550mm x (W) 710mm x (D) 720mm
Unboxed Dimensions (mm)	(H)1525mm x (W) 600mm x (D) 680mm
Weight (Kg's) - boxed	80kg
Shipping Volume (Crated)	1.04m ³

Container Quantities

20'	35
40'	73
40'HC	73

Universal Motor Manufacturing Corporation • Brits, South Africa

012 250 2703 • 086 514 6975 • Int +27 12 250 2730 • Int +27 86 514 6975

info@ummac.co.za • www.ummac.co.za

E&OE Errors and Omissions Excepted